

**ALCOHOL INFLUENCE REPORT FORM, ALCOTEST 7110 MKIII-C
FAIRVIEW POLICE**

Department Case No.: 13-14437
 Summons No(s): D137171
 Sequential File No.: 00433

Subject

Last Name: ALVAREZ First Name: MANUEL MI: .
 D.O.B.: 03/22/1992 Age: 21 Gender: MALE Ht: 6 ft. 00 in. Wt: 270 lbs.
 Driver License Number: [REDACTED] Issuing State: NJ

Arresting Officer

Last Name: BUCCO First Name: JOSEPH MI: .
 Badge No.: 24 Arrest Date: 12/08/2013 Arrest Time: 01:30S Arrest Location: 0218

Instrument


Alcotest 7110 MKIII-C Serial No.: ARXE-0081
 Location: FAIRVIEW POLICE
 Calibration File No.: 00419 Calib. Date: 09/30/2013 Calib. No.: 00013
 Certification File No.: 00420 Cert. Date: 09/30/2013 Cert. No.: 00011
 Linearity File No.: 00421 Lin. Date: 09/30/2013 Lin. No.: 00011
 Solution File No.: 00432 Soln. Date: 12/08/2013 Soln. No.: 00109
 Sequential File No.: 00433 File Date: 12/08/2013
 Calibrating Unit: WET Model No.: CU-34 Serial No.: DDZL-0155
 Control Solution %: 0.100% Expires: 04/19/2015
 Solution Control Lot: 13D112 Bottle No.: 0209

Breath Test Information

Function	Result %BAC	Time HH:MM	Volume (L)	Duration Sec (s)	Date of Test: 12/08/2013 Temp. Sim.(°C)	Error Message
Ambient Air Blank	0.000%	02:23S				
Control Test 1					34.0°C	
EC Result	0.100%	02:23S				
IR Result	0.099%	02:23S				
Ambient Air Blank	0.000%	02:24S				
Breath Test 1			3.9L	12.2s		
EC Result	0.000%	02:25S				
IR Result	0.000%	02:25S				
Ambient Air Blank	0.000%	02:25S				
Breath Test 2			3.6L	13.4s		
EC Result	0.000%	02:28S				
IR Result	0.000%	02:28S				
Ambient Air Blank	0.000%	02:29S				
Control Test 2					34.0°C	
EC Result	0.100%	02:29S				
IR Result	0.098%	02:29S				
Ambient Air Blank	0.000%	02:30S				

REPORTED BREATH TEST RESULT: 0.00% BAC

Breath Test Operator

Last Name: MARTIC First Name: MICHAEL MI: J
 Signature:  #18 Badge No.: 18
 Date: 12/08/2013

Copy Given to Subject